
Midmark M11 - v1.0.5
Total Cycles: 2534

Name: RITTER2
Sterilizer ID: V2313228
06/19/2023 04:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.8 F 0.0
2:00 174.3 F 3.3
4:00 184.6 F 7.5
6:00 207.3 F 12.2
8:00 230.0 F 7.2
10:00 239.6 F 12.2
12:00 237.9 F 9.7
14:00 236.5 F 9.5
16:00 248.6 F 14.6
18:00 259.8 F 20.9
20:00 269.9 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.4 F 29.5
2:00 272.4 F 29.5
3:00 272.4 F 29.6
4:00 271.9 F 29.7

Min 270.6 F 28.5
Max 273.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 223.3 F 0.0
10:00 236.3 F 0.0
15:00 227.7 F 0.0
20:00 226.5 F 0.0
25:00 228.3 F 0.0
30:00 229.9 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:00:22
CYCLE COMPLETE

06/19/2023 05:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____