
Midmark M11 - v1.0.5
Total Cycles: 448

Name: RITTER2
Sterilizer ID: V2313228
10/14/2021 07:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.0 F 0.0
2:00 177.6 F 2.0
4:00 194.5 F 6.8
6:00 214.7 F 12.7
8:00 230.1 F 8.7
10:00 236.2 F 10.2
12:00 243.4 F 13.1
14:00 243.7 F 12.5
16:00 256.8 F 19.6
18:00 268.5 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.3 F 29.2
2:00 271.9 F 29.5
3:00 271.9 F 29.0
4:00 272.3 F 29.7

Min 271.4 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 229.0 F 0.0
10:00 216.5 F 0.0
15:00 224.4 F 0.0
20:00 216.0 F 0.0
25:00 215.7 F 0.0
30:00 212.6 F 0.0
Drying Complete

FILLING: 2:27
HEATING: 19:05
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

10/14/2021 08:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____