

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1955

Name: RITTER 1  
Sterilizer ID: V2178198  
05/26/2022 08:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.8 F 0.0  
2:00 136.3 F 1.2  
4:00 147.3 F 4.0  
6:00 172.0 F 7.7  
8:00 190.8 F 12.1  
10:00 213.5 F 5.3  
12:00 226.6 F 9.8  
14:00 231.9 F 8.6  
16:00 238.6 F 11.1  
18:00 235.2 F 8.4  
20:00 248.8 F 14.8  
22:00 259.9 F 21.0  
24:00 270.3 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.5 F 29.2  
2:00 271.2 F 28.4  
3:00 272.1 F 29.2  
4:00 272.1 F 28.9

Min 271.0 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.9 F 0.0  
5:00 215.0 F 0.0  
10:00 210.6 F 0.0  
15:00 181.1 F 0.0  
20:00 173.9 F 0.0  
25:00 159.7 F 0.0  
30:00 152.9 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 24:54  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:03:07  
CYCLE COMPLETE

05/26/2022 09:45 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02