
Midmark M11 - v1.0.0
Total Cycles: 5024

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2024 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.4 F 0.0
2:00 167.0 F 0.3
4:00 176.9 F 1.2
6:00 196.7 F 5.3
8:00 211.0 F 9.5
10:00 230.0 F 10.1
12:00 229.3 F 8.3
14:00 240.4 F 13.1
16:00 239.9 F 10.4
18:00 236.5 F 8.3
20:00 249.6 F 14.7
22:00 260.8 F 21.0
24:00 270.9 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.9
2:00 271.6 F 28.2
3:00 272.7 F 29.0
4:00 272.1 F 28.5

Min 271.5 F 28.2
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 247.1 F 0.0
10:00 211.1 F 0.0
15:00 204.7 F 0.0
20:00 216.9 F 0.0
25:00 231.4 F 0.0
30:00 232.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:36
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:04:14
CYCLE COMPLETE

10/25/2024 10:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____