
Midmark M11 - v1.0.0
Total Cycles: 3240

Name: RITTER 1
Sterilizer ID: V2178198
05/16/2023 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 226.4 F 0.0
2:00 217.4 F 3.5
4:00 222.8 F 9.2
6:00 229.2 F 7.7
8:00 237.5 F 11.1
10:00 236.3 F 9.0
12:00 239.1 F 9.4
14:00 249.4 F 15.1
16:00 261.9 F 22.2
18:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.1
2:00 271.5 F 28.6
3:00 272.6 F 29.2
4:00 271.8 F 28.6

Min 271.5 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.9 F 0.0
5:00 231.5 F 0.0
10:00 231.7 F 0.0
15:00 227.5 F 0.0
20:00 223.8 F 0.0
25:00 219.2 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:17
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:57:48
CYCLE COMPLETE

05/16/2023 11:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____