
Midmark M11 - v1.0.5
Total Cycles: 867

Name: RITTER2
Sterilizer ID: V2313228
03/09/2022 01:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.0 F 0.0
2:00 149.0 F 2.5
4:00 171.5 F 6.1
6:00 195.4 F 11.0
8:00 214.4 F 4.9
10:00 225.5 F 10.4
12:00 229.8 F 7.6
14:00 242.8 F 13.4
16:00 242.8 F 12.3
18:00 255.6 F 19.0
20:00 266.9 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.6 F 29.9
2:00 271.5 F 29.2
3:00 271.5 F 29.7
4:00 273.3 F 30.5

Min 270.9 F 29.0
Max 273.5 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 214.1 F 0.0
10:00 202.3 F 0.0
15:00 206.9 F 0.0
20:00 204.0 F 0.0
25:00 203.5 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:07
CYCLE COMPLETE

03/09/2022 02:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____