

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1192

Name: RITTER2  
Sterilizer ID: V2313228  
06/16/2022 06:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.7 F 0.0  
2:00 73.9 F 0.1  
4:00 79.0 F 0.8  
6:00 103.0 F 2.7  
8:00 133.9 F 5.5  
10:00 164.0 F 8.9  
12:00 185.6 F 13.1  
14:00 203.4 F 4.9  
16:00 222.0 F 8.5  
18:00 231.9 F 12.3  
20:00 231.3 F 8.3  
22:00 242.6 F 13.3  
24:00 240.8 F 11.3  
26:00 251.7 F 17.0  
28:00 262.3 F 23.2  
30:00 272.5 F 30.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.8 F 29.9  
2:00 271.4 F 29.4  
3:00 272.9 F 30.1  
4:00 271.7 F 29.4  
  
Min 271.3 F 29.3  
Max 273.2 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.4 F 0.0  
5:00 227.6 F 0.0  
10:00 230.7 F 0.0  
15:00 191.9 F 0.0  
20:00 180.7 F 0.0  
25:00 181.7 F 0.0  
30:00 181.2 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 30:26  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:08:32  
CYCLE COMPLETE

06/16/2022 07:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_