

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4135

Name: RITTER2  
Sterilizer ID: V2313228  
11/19/2024 11:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.8 F 0.0  
2:00 150.1 F 0.0  
4:00 164.3 F 1.5  
6:00 188.2 F 5.0  
8:00 209.3 F 8.7  
10:00 217.2 F 13.0  
12:00 225.9 F 7.9  
14:00 237.4 F 12.7  
16:00 236.4 F 10.1  
18:00 238.1 F 9.3  
20:00 247.7 F 14.6  
22:00 258.7 F 20.7  
24:00 269.0 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.8 F 29.8  
2:00 271.8 F 29.2  
3:00 271.2 F 29.3  
4:00 272.4 F 29.8

Min 271.0 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 204.2 F 0.0  
10:00 194.8 F 0.0  
15:00 191.4 F 0.0  
20:00 187.2 F 0.0  
25:00 184.0 F 0.0  
30:00 181.9 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:11  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:03:54  
CYCLE COMPLETE

11/19/2024 12:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_