

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5006

Name: RITTER 1  
Sterilizer ID: V2178198  
10/17/2024 01:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 114.8 F 0.0  
2:00 118.1 F 0.2  
4:00 145.5 F 0.6  
6:00 174.6 F 2.1  
8:00 194.7 F 5.5  
10:00 223.8 F 9.5  
12:00 227.0 F 11.2  
14:00 230.2 F 7.7  
16:00 240.3 F 12.1  
18:00 237.2 F 9.1  
20:00 241.1 F 10.1  
22:00 247.1 F 13.4  
24:00 258.5 F 19.6  
26:00 268.9 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 27.9  
1:00 272.5 F 28.7  
2:00 271.9 F 28.1  
3:00 271.8 F 28.6  
4:00 272.5 F 28.7

Min 271.0 F 27.8  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.6 F 0.0  
5:00 227.5 F 0.0  
10:00 232.5 F 0.0  
15:00 230.8 F 0.0  
20:00 221.5 F 0.0  
25:00 224.1 F 0.0  
30:00 220.3 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 27:12  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:05:49  
CYCLE COMPLETE

10/17/2024 02:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_