

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3070

Name: RITTER2  
Sterilizer ID: V2313228  
11/01/2023 11:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.2 F 0.0  
2:00 168.0 F 2.1  
4:00 183.4 F 6.7  
6:00 203.9 F 11.4  
8:00 223.4 F 6.3  
10:00 228.7 F 11.6  
12:00 234.5 F 9.2  
14:00 238.7 F 9.9  
16:00 248.6 F 14.8  
18:00 260.8 F 21.7  
20:00 271.9 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.9  
1:00 272.4 F 29.4  
2:00 271.0 F 28.8  
3:00 272.9 F 30.1  
4:00 272.0 F 29.3

Min 270.9 F 28.7  
Max 273.2 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.0 F 0.0  
5:00 211.2 F 0.0  
10:00 198.3 F 0.0  
15:00 197.0 F 0.0  
20:00 195.0 F 0.0  
25:00 193.8 F 0.0  
30:00 193.0 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 20:33  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 00:59:43  
CYCLE COMPLETE

11/01/2023 12:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_