
Midmark M11 - v1.0.0
Total Cycles: 2540

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2022 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.5 F 0.0
2:00 185.4 F 4.3
4:00 189.0 F 8.8
6:00 209.0 F 9.7
8:00 223.8 F 7.7
10:00 237.0 F 13.0
12:00 237.3 F 10.1
14:00 238.4 F 8.2
16:00 248.9 F 14.9
18:00 260.8 F 21.6
20:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.2
2:00 271.5 F 28.5
3:00 272.5 F 29.4
4:00 272.2 F 28.9

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 217.3 F 0.0
10:00 230.7 F 0.0
15:00 230.3 F 0.0
20:00 228.9 F 0.0
25:00 229.3 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 20:36
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

11/10/2022 01:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____