
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000915
10:26 07/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:18
STERILIZING 4:20
136.0 C 220kPa 4:20
136.4 C 220kPa 4:51
136.2 C 222kPa 5:21
136.4 C 221kPa 5:51
136.4 C 225kPa 6:21
136.3 C 224kPa 6:51
136.3 C 222kPa 7:21
136.4 C 222kPa 7:51
136.4 C 222kPa 7:51
VENTING 7:51
AIR DRYING 8:46
CYCLE COMPLETE 8:49

Digital Signature #
C9824B96B565DADB
