
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003115
14:30 08/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:28
STERILIZING 3:28
136.4 C 220kPa 3:28
136.4 C 226kPa 3:58
136.4 C 224kPa 4:28
136.4 C 225kPa 4:58
136.4 C 225kPa 5:28
136.4 C 225kPa 5:58
136.4 C 221kPa 6:28
136.2 C 220kPa 6:58
136.2 C 220kPa 6:58
VENTING 6:58
AIR DRYING 7:52
CYCLE COMPLETE 7:55

Digital Signature #
2CE11A98ED5B50B9
