
Midmark M11 - v1.0.5
Total Cycles: 1369

Name: RITTER2
Sterilizer ID: V2313228
08/04/2022 11:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.5 F 0.0
2:00 150.0 F 0.0
4:00 169.5 F 0.7
6:00 193.3 F 5.0
8:00 217.3 F 9.8
10:00 226.2 F 8.0
12:00 235.3 F 11.0
14:00 234.4 F 8.6
16:00 240.3 F 10.5
18:00 247.4 F 14.3
20:00 259.4 F 21.0
22:00 270.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.7
2:00 272.0 F 29.0
3:00 271.7 F 29.3
4:00 272.6 F 29.7

Min 271.4 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 233.9 F 0.0
10:00 221.6 F 0.0
15:00 230.4 F 0.0
20:00 228.9 F 0.0
25:00 229.3 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 22:54
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:12
CYCLE COMPLETE

08/04/2022 12:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____