
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005290
12:02 28/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:33
STERILIZING 3:28
136.0 C 220kPa 3:28
136.4 C 225kPa 3:58
136.3 C 225kPa 4:28
136.3 C 225kPa 4:58
136.4 C 223kPa 5:28
136.4 C 223kPa 5:58
136.2 C 223kPa 6:28
136.4 C 224kPa 6:58
136.4 C 224kPa 6:58
VENTING 6:59
AIR DRYING 8:55
CYCLE COMPLETE 9:08

Digital Signature #
18B3141E2847F7A4
