
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000802
14:47 17/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:56
PRESSURIZING 6:49
STERILIZING 8:30
136.5 C 223kPa 8:30
136.4 C 224kPa 9:00
136.4 C 223kPa 9:30
136.4 C 219kPa 10:00
135.7 C 218kPa 10:30
136.3 C 222kPa 11:00
136.4 C 221kPa 11:30
136.1 C 221kPa 12:00
136.1 C 221kPa 12:00
VENTING 12:00
AIR DRYING 13:03
CYCLE COMPLETE 37:23

Digital Signature #
DF264D568B436835
