
Midmark M11 - v1.0.0
Total Cycles: 3929

Name: RITTER 1
Sterilizer ID: V2178198
10/20/2023 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.4 F 0.0
2:00 156.0 F 2.5
4:00 169.0 F 5.7
6:00 216.9 F 9.7
8:00 217.3 F 5.4
10:00 221.9 F 7.8
12:00 234.8 F 12.6
14:00 236.5 F 9.4
16:00 238.3 F 9.6
18:00 246.6 F 13.6
20:00 258.5 F 19.9
22:00 268.9 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.3 F 28.9
2:00 272.7 F 29.3
3:00 271.5 F 28.7
4:00 272.4 F 29.1

Min 271.5 F 28.6
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.5 F 0.0
5:00 232.6 F 0.0
10:00 219.6 F 0.0
15:00 238.4 F 0.0
20:00 233.0 F 0.0
25:00 212.3 F 0.0
30:00 215.6 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:52
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:32
CYCLE COMPLETE

10/20/2023 12:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____