

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000198  
8:41 21/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:04  
PRESSURIZING 4:05  
STERILIZING 5:45  
136.4 C 233kPa 5:45  
136.3 C 232kPa 6:15  
136.4 C 233kPa 6:45  
136.4 C 233kPa 7:15  
136.4 C 233kPa 7:45  
136.4 C 230kPa 8:15  
136.2 C 232kPa 8:45  
136.3 C 232kPa 9:15  
136.3 C 232kPa 9:15  
VENTING 9:15  
AIR DRYING 10:17  
CYCLE COMPLETE 10:19  
-----

Digital Signature #  
CDD170A150A874DB  
-----