

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1990

Name: RITTER2  
Sterilizer ID: V2313228  
02/01/2023 08:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.6 F 0.0  
2:00 79.8 F 0.0  
4:00 104.7 F 0.0  
6:00 147.6 F 0.5  
8:00 152.3 F 3.4  
10:00 180.1 F 6.4  
12:00 196.5 F 10.3  
14:00 220.2 F 6.1  
16:00 223.5 F 7.4  
18:00 234.1 F 11.7  
20:00 233.8 F 8.2  
22:00 243.0 F 12.1  
24:00 243.8 F 12.6  
26:00 256.4 F 19.2  
28:00 267.6 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.4 F 29.3  
2:00 271.8 F 29.5  
3:00 272.0 F 29.1  
4:00 272.3 F 29.7

Min 271.2 F 28.8  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.3 F 0.0  
5:00 200.2 F 0.0  
10:00 204.8 F 0.0  
15:00 206.4 F 0.0  
20:00 210.2 F 0.0  
25:00 215.1 F 0.0  
30:00 213.8 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 29:19  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:07:17  
CYCLE COMPLETE

02/01/2023 09:15 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_