

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 003710  
11:57 27/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:56  
PRESSURIZING 2:02  
STERILIZING 2:57  
136.4 C 220kPa 2:57  
136.1 C 224kPa 3:27  
136.4 C 223kPa 3:57  
136.5 C 224kPa 4:27  
136.0 C 221kPa 4:57  
136.4 C 224kPa 5:27  
136.2 C 222kPa 5:57  
136.4 C 223kPa 6:27  
136.4 C 223kPa 6:27  
VENTING 6:27  
AIR DRYING 7:18  
CYCLE COMPLETE 7:27

-----  
Digital Signature #  
BBB12E0E903F6A6C  
-----