

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1221

Name: RITTER2  
Sterilizer ID: V2313228  
06/24/2022 01:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.4 F 0.0  
2:00 193.3 F 3.6  
4:00 196.2 F 8.2  
6:00 213.5 F 13.5  
8:00 225.4 F 8.6  
10:00 236.0 F 11.3  
12:00 240.3 F 11.9  
14:00 240.0 F 10.8  
16:00 252.8 F 17.3  
18:00 264.2 F 23.9  
20:00 271.5 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.8  
1:00 272.2 F 29.7  
2:00 272.3 F 29.3  
3:00 271.4 F 28.8  
4:00 271.2 F 29.1  
  
Min 271.0 F 28.6  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.4 F 0.0  
5:00 228.1 F 0.0  
10:00 241.4 F 0.0  
15:00 230.9 F 0.0  
20:00 223.7 F 0.0  
25:00 222.6 F 0.0  
30:00 221.6 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 20:08  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

06/24/2022 02:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_