
Midmark M11 - v1.0.0
Total Cycles: 914

Name: RITTER 1
Sterilizer ID: V2178198
06/18/2021 06:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.6 F 0.0
2:00 90.4 F 0.0
4:00 128.2 F 0.0
6:00 131.7 F 0.0
8:00 152.7 F 0.0
10:00 198.2 F 0.1
12:00 196.6 F 0.3
14:00 210.7 F 2.7
16:00 220.0 F 5.0
18:00 227.3 F 7.5
20:00 234.1 F 10.6
22:00 236.3 F 9.5
24:00 239.7 F 10.6
26:00 237.9 F 9.0
28:00 246.8 F 13.5
30:00 244.3 F 12.2
32:00 255.2 F 17.8
34:00 265.4 F 23.9
36:00 271.4 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.5 F 28.8
2:00 272.3 F 28.4
3:00 271.5 F 27.9
4:00 272.1 F 28.6

Min 271.4 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 233.8 F 0.0
10:00 218.6 F 0.0
15:00 208.9 F 0.0
20:00 199.4 F 0.0
25:00 221.2 F 0.0
30:00 232.0 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 36:01
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:14:28
CYCLE COMPLETE

06/18/2021 07:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____