
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 001524
13:56 10/21/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:21
PRESSURIZING 6:12
STERILIZING 7:45
136.0 C 220kPa 7:45
136.5 C 218kPa 8:15
136.4 C 223kPa 8:45
136.4 C 221kPa 9:15
136.3 C 221kPa 9:45
136.3 C 222kPa 10:15
136.2 C 221kPa 10:45
136.2 C 221kPa 11:15
VENTING 11:15
AIR DRYING 12:19
CYCLE COMPLETE 42:19

Digital Signature #
2500870B733956B3
