
Midmark M11 - v1.0.0
Total Cycles: 745

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2021 06:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 71.9 F 0.0
4:00 77.7 F 0.0
6:00 107.1 F 0.1
8:00 148.6 F 0.2
10:00 185.2 F 0.3
12:00 204.1 F 3.0
14:00 219.4 F 5.9
16:00 228.1 F 9.0
18:00 235.9 F 12.8
20:00 233.8 F 9.4
22:00 239.0 F 9.5
24:00 244.3 F 13.0
26:00 243.2 F 11.6
28:00 255.9 F 18.2
30:00 267.2 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.4 F 28.8
2:00 271.4 F 28.2
3:00 272.2 F 28.7
4:00 271.3 F 28.1

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 227.9 F 0.0
10:00 207.2 F 0.0
15:00 215.2 F 0.0
20:00 205.9 F 0.0
25:00 197.2 F 0.0
30:00 195.1 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:26
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:09:11
CYCLE COMPLETE

03/25/2021 08:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____