

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3973

Name: RITTER 1  
Sterilizer ID: V2178198  
11/02/2023 12:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.7 F 0.0  
2:00 182.1 F 4.1  
4:00 201.1 F 9.4  
6:00 221.3 F 7.4  
8:00 231.9 F 10.6  
10:00 231.9 F 7.6  
12:00 242.3 F 11.6  
14:00 245.3 F 13.0  
16:00 258.0 F 19.7  
18:00 269.1 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.2 F 29.0  
2:00 272.4 F 28.8  
3:00 271.8 F 28.4  
4:00 271.2 F 28.1  
  
Min 271.1 F 28.0  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 209.6 F 0.0  
10:00 216.6 F 0.0  
15:00 219.8 F 0.0  
20:00 222.7 F 0.0  
25:00 225.2 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 2:07  
HEATING: 19:09  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 00:58:43  
CYCLE COMPLETE

11/02/2023 01:01 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_