
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006208
9:49 30/08/2023

WRAPPED

135 C for 10 min.

CYCLE START	0:00
CONDITIONING	1:51
PRESSURIZING	3:49
STERILIZING	4:39
136.4 C 222kPa	4:39
136.3 C 223kPa	5:09
136.3 C 223kPa	5:39
136.3 C 224kPa	6:09
136.0 C 217kPa	6:39
136.4 C 221kPa	7:09
136.4 C 221kPa	7:39
136.4 C 219kPa	8:09
136.4 C 224kPa	8:39
136.4 C 222kPa	9:09
136.3 C 225kPa	9:39
136.4 C 223kPa	10:09
136.2 C 222kPa	10:39
136.1 C 220kPa	11:09
136.4 C 223kPa	11:39
136.4 C 224kPa	12:09
136.0 C 218kPa	12:39
136.4 C 223kPa	13:09
136.2 C 217kPa	13:39
136.4 C 224kPa	14:09
136.4 C 220kPa	14:39
136.4 C 220kPa	14:39
VENTING	14:39
AIR DRYING	15:33
CYCLE COMPLETE	33:10

Digital Signature #

10397521F7428F8B
