
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001577
15:45 08/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:12
STERILIZING 3:11
136.4 C 224kPa 3:11
136.1 C 222kPa 3:41
136.4 C 226kPa 4:11
136.3 C 225kPa 4:41
136.0 C 223kPa 5:11
136.2 C 224kPa 5:41
136.3 C 224kPa 6:11
136.4 C 225kPa 6:41
136.4 C 225kPa 6:41
VENTING 6:41
AIR DRYING 7:32
CYCLE COMPLETE 8:57

Digital Signature #
FB5338F403D561C8
