
Midmark M11 - v1.0.0
Total Cycles: 974

Name: RITTER 1
Sterilizer ID: V2178198
07/07/2021 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.2 F 0.0
2:00 170.1 F 2.3
4:00 175.5 F 6.1
6:00 195.8 F 10.8
8:00 223.7 F 4.5
10:00 226.1 F 10.1
12:00 233.5 F 7.9
14:00 244.8 F 12.7
16:00 241.9 F 10.9
18:00 254.4 F 17.2
20:00 265.1 F 23.7
22:00 270.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.5 F 28.6
2:00 271.5 F 28.6
3:00 271.3 F 28.5
4:00 271.5 F 28.4

Min 271.3 F 28.4
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.9 F 0.0
5:00 203.2 F 0.0
10:00 205.2 F 0.0
15:00 214.5 F 0.0
20:00 225.0 F 0.0
25:00 244.8 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 22:32
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:42
CYCLE COMPLETE

07/07/2021 11:03 AM
ACCEPTED: X
REJECTED:
Operator: U01