

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4171

Name: RITTER2  
Sterilizer ID: V2313228  
11/29/2024 04:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.3 F 0.0  
2:00 177.1 F 1.1  
4:00 186.8 F 5.3  
6:00 207.8 F 10.2  
8:00 221.6 F 6.4  
10:00 233.1 F 10.3  
12:00 233.4 F 8.5  
14:00 242.1 F 11.8  
16:00 246.3 F 13.9  
18:00 258.9 F 20.8  
20:00 270.2 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.7 F 29.7  
2:00 271.7 F 28.9  
3:00 271.9 F 29.6  
4:00 272.5 F 29.6  
  
Min 271.3 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.6 F 0.0  
5:00 237.9 F 0.0  
10:00 218.1 F 0.0  
15:00 203.2 F 0.0  
20:00 196.8 F 0.0  
25:00 204.9 F 0.0  
30:00 212.5 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 20:54  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 00:59:59  
CYCLE COMPLETE

11/29/2024 05:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_