
Midmark M11 - v1.0.0
Total Cycles: 3912

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2023 12:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.3 F 0.0
2:00 139.9 F 2.0
4:00 158.0 F 5.3
6:00 185.9 F 9.7
8:00 213.4 F 6.2
10:00 221.3 F 8.2
12:00 233.8 F 13.2
14:00 236.0 F 9.9
16:00 236.3 F 8.6
18:00 247.6 F 14.3
20:00 259.8 F 21.0
22:00 270.9 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.4 F 29.1
2:00 271.4 F 28.7
3:00 272.5 F 29.2
4:00 271.4 F 28.5

Min 271.3 F 28.5
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 212.8 F 0.0
10:00 223.2 F 0.0
15:00 220.3 F 0.0
20:00 219.1 F 0.0
25:00 219.7 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:45
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:01:13
CYCLE COMPLETE

10/17/2023 01:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____