
Midmark M11 - v1.0.5
Total Cycles: 3375

Name: RITTER2
Sterilizer ID: V2313228
02/12/2024 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.0 F 0.0
2:00 70.9 F 0.0
4:00 104.0 F 0.0
6:00 107.4 F 1.1
8:00 143.5 F 3.8
10:00 168.3 F 7.1
12:00 190.1 F 10.7
14:00 201.2 F 5.8
16:00 215.1 F 6.4
18:00 226.1 F 10.1
20:00 232.3 F 9.5
22:00 237.9 F 11.0
24:00 235.3 F 8.7
26:00 248.6 F 15.0
28:00 259.7 F 21.2
30:00 269.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.3 F 29.6
2:00 272.2 F 29.1
3:00 272.2 F 29.5
4:00 272.5 F 29.4

Min 271.5 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 197.6 F 0.0
10:00 193.4 F 0.0
15:00 192.3 F 0.0
20:00 188.4 F 0.0
25:00 185.7 F 0.0
30:00 183.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:49
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:50
CYCLE COMPLETE

02/12/2024 10:55 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____