
Midmark M11 - v1.0.0
Total Cycles: 1615

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2022 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.1 F 0.0
2:00 149.5 F 2.5
4:00 177.1 F 6.4
6:00 214.6 F 11.2
8:00 221.2 F 5.1
10:00 230.9 F 10.2
12:00 232.3 F 7.6
14:00 243.6 F 12.9
16:00 242.2 F 11.4
18:00 255.1 F 18.0
20:00 266.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.7 F 29.3
2:00 272.3 F 28.8
3:00 271.2 F 28.3
4:00 271.6 F 28.7

Min 271.1 F 28.2
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 243.2 F 0.0
10:00 234.4 F 0.0
15:00 228.5 F 0.0
20:00 217.2 F 0.0
25:00 220.4 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:35
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

02/10/2022 11:49 AM
ACCEPTED: X
REJECTED:
Operator: U02