
Midmark M11 - v1.0.5
Total Cycles: 3078

Name: RITTER2
Sterilizer ID: V2313228
11/02/2023 01:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.4 F 0.0
2:00 114.3 F 0.5
4:00 144.6 F 2.8
6:00 171.5 F 7.1
8:00 195.6 F 12.2
10:00 213.0 F 5.1
12:00 228.1 F 9.8
14:00 232.7 F 8.8
16:00 241.5 F 12.2
18:00 240.8 F 11.0
20:00 254.5 F 17.9
22:00 266.5 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.1
2:00 272.1 F 29.2
3:00 272.3 F 28.9
4:00 272.1 F 29.2

Min 271.5 F 28.5
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.4 F 0.0
5:00 216.6 F 0.0
10:00 204.7 F 0.0
15:00 203.2 F 0.0
20:00 200.9 F 0.0
25:00 199.8 F 0.0
30:00 199.2 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 23:27
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:37
CYCLE COMPLETE

11/02/2023 02:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____