
Midmark M11 - v1.0.5
Total Cycles: 2415

Name: RITTER2
Sterilizer ID: V2313228
05/25/2023 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.1 F 0.0
2:00 190.6 F 4.2
4:00 206.4 F 9.8
6:00 215.6 F 6.5
8:00 230.5 F 11.0
10:00 233.4 F 8.8
12:00 238.4 F 10.1
14:00 248.6 F 14.7
16:00 260.8 F 21.5
18:00 271.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.4
2:00 271.9 F 28.8
3:00 271.3 F 28.9
4:00 272.3 F 29.4

Min 271.0 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.8 F 0.0
5:00 230.1 F 0.0
10:00 243.3 F 0.0
15:00 238.1 F 0.0
20:00 226.5 F 0.0
25:00 225.2 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 18:38
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

05/25/2023 12:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____