
Midmark M11 - v1.0.5
Total Cycles: 3653

Name: RITTER2
Sterilizer ID: V2313228
05/28/2024 02:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.4 F 0.0
2:00 122.8 F 0.8
4:00 131.4 F 3.2
6:00 167.8 F 6.4
8:00 183.1 F 10.1
10:00 205.5 F 8.8
12:00 216.4 F 7.1
14:00 226.2 F 11.2
16:00 227.9 F 7.3
18:00 239.8 F 12.8
20:00 239.5 F 11.2
22:00 251.5 F 17.1
24:00 262.1 F 23.3
26:00 272.0 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 271.9 F 30.2
2:00 272.3 F 30.0
3:00 271.1 F 29.7
4:00 272.7 F 30.4

Min 271.0 F 29.5
Max 273.2 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 202.0 F 0.0
10:00 197.7 F 0.0
15:00 186.8 F 0.0
20:00 184.8 F 0.0
25:00 195.1 F 0.0
30:00 198.7 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 26:25
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:05:17
CYCLE COMPLETE

05/28/2024 03:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____