
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001406
12:55 09/08/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 5:31
STERILIZING 6:48
136.0 C 220kPa 6:48
136.5 C 217kPa 7:18
136.4 C 223kPa 7:48
136.3 C 223kPa 8:18
136.4 C 223kPa 8:48
136.3 C 222kPa 9:18
136.4 C 223kPa 9:48
136.0 C 220kPa 10:18
VENTING 10:18
AIR DRYING 11:19
CYCLE COMPLETE 11:33

Digital Signature #
46E0B631FB27A252
