
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001278
16:57 28/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:24
PRESSURIZING 3:27
STERILIZING 4:35
136.3 C 220kPa 4:35
136.4 C 225kPa 5:05
136.4 C 221kPa 5:35
136.2 C 223kPa 6:05
136.4 C 225kPa 6:35
136.3 C 224kPa 7:05
136.4 C 224kPa 7:35
136.4 C 224kPa 8:05
136.4 C 224kPa 8:05
VENTING 8:05
AIR DRYING 8:56
CYCLE COMPLETE 23:21

Digital Signature #
B53E619D17DECB27
