
Midmark M11 - v1.0.5
Total Cycles: 1758

Name: RITTER2
Sterilizer ID: V2313228
11/22/2022 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.6 F 0.0
2:00 165.4 F 0.6
4:00 188.5 F 3.8
6:00 209.7 F 8.7
8:00 220.8 F 9.3
10:00 233.6 F 11.1
12:00 236.7 F 9.3
14:00 237.3 F 9.2
16:00 250.6 F 15.8
18:00 263.1 F 23.2
20:00 271.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.3 F 29.7
2:00 272.8 F 29.7
3:00 272.0 F 29.6
4:00 272.8 F 29.8

Min 271.0 F 28.7
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 217.9 F 0.0
10:00 233.6 F 0.0
15:00 220.4 F 0.0
20:00 218.9 F 0.0
25:00 218.4 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

11/22/2022 11:06 AM
ACCEPTED: X
REJECTED:
Operator: U02