

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1989

Name: RITTER2  
Sterilizer ID: V2313228  
01/31/2023 04:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.2 F 0.0  
2:00 151.6 F 2.2  
4:00 168.8 F 5.7  
6:00 189.5 F 10.5  
8:00 212.4 F 4.9  
10:00 226.5 F 9.8  
12:00 232.5 F 8.6  
14:00 241.5 F 12.8  
16:00 242.3 F 11.7  
18:00 255.1 F 18.5  
20:00 266.6 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 272.5 F 29.7  
2:00 271.4 F 29.0  
3:00 272.4 F 29.9  
4:00 271.9 F 29.4

Min 271.1 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
5:00 212.6 F 0.0  
10:00 233.5 F 0.0  
15:00 227.9 F 0.0  
20:00 227.9 F 0.0  
25:00 227.6 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 21:32  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:43  
CYCLE COMPLETE

01/31/2023 05:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_