
Midmark M11 - v1.0.0
Total Cycles: 3122

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2023 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.6 F 0.0
2:00 127.2 F 0.0
4:00 150.4 F 0.0
6:00 187.8 F 3.2
8:00 203.2 F 7.8
10:00 218.1 F 12.0
12:00 223.5 F 6.6
14:00 235.6 F 11.5
16:00 233.5 F 8.4
18:00 241.5 F 11.2
20:00 245.2 F 13.0
22:00 257.4 F 19.5
24:00 268.4 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.3 F 29.2
2:00 272.0 F 28.6
3:00 271.9 F 28.8
4:00 272.4 F 29.0

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 210.7 F 0.0
10:00 227.0 F 0.0
15:00 231.9 F 0.0
20:00 232.5 F 0.0
25:00 215.2 F 0.0
30:00 214.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 25:06
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:03:14
CYCLE COMPLETE

04/18/2023 12:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____