

-----  
Midmark M11 - v1.0.5  
Total Cycles: 242

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2021 06:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.3 F 0.0  
2:00 78.0 F 0.0  
4:00 94.9 F 0.1  
6:00 105.6 F 1.6  
8:00 135.3 F 4.6  
10:00 171.9 F 8.3  
12:00 192.0 F 12.6  
14:00 207.4 F 6.0  
16:00 219.6 F 10.0  
18:00 228.6 F 9.5  
20:00 236.9 F 11.5  
22:00 235.0 F 9.4  
24:00 248.3 F 15.5  
26:00 259.7 F 21.9  
28:00 269.2 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 30.6  
1:00 272.4 F 30.6  
2:00 272.7 F 30.8  
3:00 272.3 F 30.6  
4:00 272.6 F 30.8

Min 272.0 F 30.4  
Max 272.7 F 30.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.3 F 0.0  
5:00 212.4 F 0.0  
10:00 220.5 F 0.0  
15:00 219.0 F 0.0  
20:00 229.2 F 0.0  
25:00 228.1 F 0.0  
Dry Cycle Aborted

FILLING: 0:54  
HEATING: 29:37  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 29:22  
TOTAL CYCLE: 01:06:54

07/26/2021 08:04 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01