
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003108
10:51 08/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:52
STERILIZING 2:59
136.4 C 222kPa 2:59
136.2 C 221kPa 3:29
136.4 C 221kPa 3:59
136.4 C 223kPa 4:29
136.4 C 221kPa 4:59
136.4 C 223kPa 5:29
136.3 C 224kPa 5:59
136.4 C 224kPa 6:29
136.4 C 224kPa 6:29
VENTING 6:29
AIR DRYING 7:22
CYCLE COMPLETE 13:07

Digital Signature #
CFC8B93820225A09
