

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
2.1uS / 1.3ppm  
CYCLE NUMBER 000033  
14:33 10/08/2020  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 3:52  
STERILIZING 4:53  
136.4 C 227kPa 4:53  
136.4 C 227kPa 5:23  
136.5 C 226kPa 5:53  
136.0 C 223kPa 6:23  
136.4 C 228kPa 6:53  
135.9 C 225kPa 7:24  
136.4 C 227kPa 7:54  
136.2 C 224kPa 8:23  
VENTING 8:23  
AIR DRYING 9:14  
CYCLE COMPLETE 39:14

-----  
Digital Signature #  
2D0AF08F42B0BEDE  
-----