
Midmark M11 - v1.0.0
Total Cycles: 1372

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2021 02:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.7 F 0.0
2:00 184.8 F 3.1
4:00 201.2 F 6.6
6:00 224.5 F 11.3
8:00 223.2 F 5.5
10:00 233.0 F 10.5
12:00 233.2 F 7.8
14:00 244.2 F 13.1
16:00 242.6 F 11.7
18:00 255.1 F 18.2
20:00 266.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.1
2:00 271.4 F 28.4
3:00 272.6 F 29.3
4:00 272.1 F 28.8

Min 271.3 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 228.8 F 0.0
10:00 222.9 F 0.0
15:00 221.1 F 0.0
20:00 219.5 F 0.0
25:00 220.2 F 0.0
30:00 217.2 F 0.0
Drying Complete

FILLING: 2:22
HEATING: 21:41
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:01:23
CYCLE COMPLETE

11/15/2021 04:03 PM
ACCEPTED: X
REJECTED:
Operator: U02