

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4303

Name: RITTER2  
Sterilizer ID: V2313228  
01/28/2025 10:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 71.1 F 0.0  
4:00 78.4 F 0.3  
6:00 113.2 F 2.4  
8:00 143.0 F 5.1  
10:00 167.6 F 8.4  
12:00 186.7 F 11.9  
14:00 205.7 F 3.3  
16:00 216.5 F 6.5  
18:00 226.6 F 10.1  
20:00 230.4 F 9.9  
22:00 236.9 F 10.7  
24:00 235.1 F 8.3  
26:00 247.3 F 14.5  
28:00 258.3 F 20.6  
30:00 268.4 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.9 F 29.8  
2:00 271.7 F 29.2  
3:00 272.9 F 29.9  
4:00 271.9 F 29.1  
  
Min 271.6 F 29.0  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.3 F 0.0  
5:00 183.6 F 0.0  
10:00 191.2 F 0.0  
15:00 180.4 F 0.0  
20:00 198.0 F 0.0  
25:00 228.1 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 31:16  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:09:40  
CYCLE COMPLETE

01/28/2025 11:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_