

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001737  
16:24 15/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:17  
PRESSURIZING 2:27  
STERILIZING 3:23  
136.0 C 220kPa 3:23  
136.4 C 224kPa 3:53  
136.3 C 224kPa 4:23  
136.4 C 223kPa 4:53  
136.2 C 225kPa 5:23  
136.4 C 226kPa 5:53  
136.4 C 226kPa 6:23  
136.3 C 225kPa 6:53  
136.3 C 225kPa 6:53  
VENTING 6:53  
AIR DRYING 7:44  
CYCLE COMPLETE 7:48  
-----

Digital Signature #  
D8937DC7C8D09AB0  
-----