
Midmark M11 - v1.0.0
Total Cycles: 1882

Name: RITTER 1
Sterilizer ID: V2178198
04/28/2022 07:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 76.9 F 0.0
4:00 92.0 F 0.0
6:00 116.1 F 0.5
8:00 142.9 F 2.2
10:00 171.5 F 4.5
12:00 191.9 F 7.5
14:00 208.0 F 10.9
16:00 215.1 F 5.3
18:00 223.0 F 7.1
20:00 233.8 F 11.2
22:00 232.9 F 6.9
24:00 243.9 F 12.8
26:00 241.8 F 10.9
28:00 254.2 F 17.3
30:00 265.2 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.3 F 28.5
2:00 271.9 F 28.5
3:00 272.0 F 28.3
4:00 272.2 F 28.8

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 189.4 F 0.0
10:00 190.2 F 0.0
15:00 203.5 F 0.0
20:00 209.1 F 0.0
25:00 215.1 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:51
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:43
CYCLE COMPLETE

04/28/2022 08:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____