
Midmark M11 - v1.0.5
Total Cycles: 3036

Name: RITTER2
Sterilizer ID: V2313228
10/20/2023 11:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.4 F 0.0
2:00 134.6 F 1.4
4:00 157.2 F 4.2
6:00 180.6 F 8.2
8:00 198.3 F 12.9
10:00 218.3 F 7.4
12:00 228.8 F 12.5
14:00 234.4 F 9.4
16:00 237.7 F 9.6
18:00 246.7 F 14.0
20:00 258.2 F 20.4
22:00 268.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.6
1:00 271.6 F 29.5
2:00 272.2 F 29.7
3:00 271.9 F 30.0
4:00 271.6 F 30.1

Min 271.1 F 29.5
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.3 F 0.0
5:00 213.3 F 0.0
10:00 193.0 F 0.0
15:00 193.7 F 0.0
20:00 192.0 F 0.0
25:00 190.1 F 0.0
30:00 188.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:52
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:02:23
CYCLE COMPLETE

10/20/2023 12:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____