
Midmark M11 - v1.0.5
Total Cycles: 3393

Name: RITTER2
Sterilizer ID: V2313228
02/16/2024 03:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.8 F 0.0
2:00 134.0 F 0.9
4:00 152.7 F 3.5
6:00 178.0 F 7.2
8:00 197.3 F 11.4
10:00 212.8 F 4.9
12:00 222.8 F 9.5
14:00 231.9 F 9.7
16:00 239.6 F 12.2
18:00 240.2 F 11.2
20:00 253.0 F 17.8
22:00 264.7 F 24.8
24:00 271.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.5
1:00 273.0 F 29.8
2:00 272.4 F 30.1
3:00 271.6 F 29.3
4:00 272.1 F 29.8

Min 271.5 F 29.3
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.7 F 0.0
5:00 199.9 F 0.0
10:00 197.4 F 0.0
15:00 197.3 F 0.0
20:00 196.7 F 0.0
25:00 195.9 F 0.0
30:00 195.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:27
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:02:58
CYCLE COMPLETE

02/16/2024 05:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____