

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 007945  
13:40 08/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:09  
STERILIZING 4:14  
136.4 C 224kPa 4:14  
136.3 C 224kPa 4:44  
136.3 C 222kPa 5:14  
136.2 C 223kPa 5:44  
136.4 C 223kPa 6:14  
136.4 C 221kPa 6:44  
136.2 C 223kPa 7:14  
136.1 C 219kPa 7:44  
136.1 C 219kPa 7:44  
VENTING 7:44  
AIR DRYING 8:40  
CYCLE COMPLETE 8:46

-----  
Digital Signature #  
EB730386D9B2C512  
-----